

IV-E244 Determination Of The Refractive Index Of Hollow Prism Using Spectrometer And Mercury Light

Scope of Learning :

- DETERMINATION OF THE REFRACTIVE INDEX OF HOLLOW PRISM USING SPECTROMETER AND MERCURY LIGHT

About Spectrometer :

- A Basic measuring instrument for quantitative spectroscopy experiments, this instrument is mounted on a stable cast iron base with attached collimator and a rotating telescope platform and graduated circle.
- A 2.4cm. Diameter achromatic optical system with a 10x Ramsden eyepiece & cross line graticule.
- An 8 cm. Diameter prism/grating Table, fully adjustable with prism & grating holders and engraved ring pattern. Protected Circle 15.0 cm with S.S. Scale.
- Vernier reading 1 minute.
- Supplied in a wooden case with 2" Brass spirit level.

Required Accessories :

Apparatus Supply:

- Spectrometer
- EDF Prism 32x32
- Hollow Glass Prism 50x50
- Spirit Level 2"
- Working Manual

Accessories (Light Source):

- Mercury Vapour Lamp Assembly
- Mercury Vapour Lamp 80 Watts
- Mercury Vapour Lamp Box
- Mercury Vapour Lamp Transformer 80 Watts

